



LAWRENCE
LIVERMORE
NATIONAL
LABORATORY

Using Non-Invasive Multi-Spectral Imaging to Quantitatively Assess Tissue Vasculature

A. Vogel, V. Chernomordik, J. Riley, M. Hassan, F. Amyot, B. Dasgeb, S. G. Demos, R. Pursley, R. Little, R. Yarchoan, Y. Tao, A. H. Gandjbakhche

October 12, 2007

Journal of Biomedical Optics

Disclaimer

This document was prepared as an account of work sponsored by an agency of the United States government. Neither the United States government nor Lawrence Livermore National Security, LLC, nor any of their employees makes any warranty, expressed or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States government or Lawrence Livermore National Security, LLC. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States government or Lawrence Livermore National Security, LLC, and shall not be used for advertising or product endorsement purposes.

Using Non-Invasive Multi-Spectral Imaging to Quantitatively Assess Tissue Vasculature

Abby Vogel^{*1,2}, Victor Chernomordik¹, Jason Riley¹, Moinuddin Hassan¹, Franck Amyot¹,
Bahar Dasgeb¹, Stavros G. Demos³, Randall Pursley⁴, Richard Little⁵, Robert Yarchoan⁵, Yang
Tao², and Amir H. Gandjbakhche¹

¹ Laboratory of Integrative and Medical Biophysics, National Institute of Child Health and
Human Development, National Institutes of Health, Bethesda, MD

² Bioimaging and Machine Vision Laboratory, Fischell Department of Bioengineering,
University of Maryland, College Park, MD

³ Lawrence Livermore National Laboratory, Livermore, CA

⁴ Center for Information Technology, National Institutes of Health, Bethesda, MD

⁵ HIV and AIDS Malignancy Branch, National Cancer Institute, National Institutes of Health,
Bethesda, MD

* Corresponding Author:

Abby Vogel

National Institutes of Health/9000 Rockville Pike

Building 9 Room B1E05

Bethesda, Md. 20892

(Office) 301-402-0648

(Fax) 301-480-2427

vogelab@mail.nih.gov

ABSTRACT:

This research describes a non-invasive, non-contact method used to quantitatively analyze the functional characteristics of tissue. Multi-spectral images collected at several near-infrared wavelengths are input into a mathematical optical skin model that considers the contributions from different analytes in the epidermis and dermis skin layers. Through a reconstruction algorithm, we can quantify the percent of blood in a given area of tissue and the fraction of that blood that is oxygenated. Imaging normal tissue confirms previously reported values for the percent of blood in tissue and the percent of blood that is oxygenated in tissue and surrounding vasculature, for the normal state and when ischemia is induced. This methodology has been applied to assess vascular Kaposi's sarcoma lesions and the surrounding tissue before and during experimental therapies. The multi-spectral imaging technique has been combined with laser Doppler imaging to gain additional information. Results indicate that these techniques are able to provide quantitative and functional information about tissue changes during experimental drug therapy and investigate progression of disease before changes are visibly apparent, suggesting a potential for them to be used as complementary imaging techniques to clinical assessment.

Keywords: multi-spectral imaging, tissue oxygenation, blood volume, drug therapy monitoring, quantitative assessment, functional information

INTRODUCTION

Quantitative instrumental methods offer a potentially objective means of assessing skin health that can supplement the conventional visual clinical observations. Moreover, such approaches can be used to provide early markers for tumor responses and to learn about the pathophysiology of the disease and its changes in response to treatment. Near-infrared spectroscopy (NIRS) is a non-contact and noninvasive method of monitoring changes in tissue blood volume and the fraction of HbO₂ and Hb.

NIRS interrogates blood in the capillaries, arterioles, and venules of tissue and enables direct and noninvasive measurements of the tissue oxygen saturation by examining absorption differences in the spectra of oxygen carriers, such as hemoglobin, myoglobin, and cytochrome aa3 (Taylor and Simonson, 1996). The concentrations of these analytes, such as HbO₂ and Hb, in tissue may be used to diagnose disease and to monitor changes in various conditions (Kollias, Gillies et al., 1995; Wagnieres, Cheng et al., 1997; Shah, Cerussi et al., 2001).

Assessing these analytes is complicated by other pigments in the skin, i.e. melanin and hemosiderin, but it is possible to correct for such pigments with NIRS. This research will address these limitations and present a method that describes quantitative changes in functional parameters. Instrumentation and techniques for spectroscopic imaging have been developed to visualize skin hemodynamics and tissue water content to test tissue viability of skin flaps (Stranc, Sowa et al., 1998; Attas, Hewko et al., 2001). Port-wine stains have been studied with NIRS because they are characterized by an abnormal density of blood vessels in the upper dermis (Svaasand, Norvang et al., 1995). NIRS has been used to follow port-wine stains through laser treatment that destroys the vessels to the extent necessary to obtain normal skin coloration.

Ryatt et al. used reflectance spectrophotometry to obtain indices of hemoglobin and melanin content of psoriatic lesions and adjacent clinically normal skin. Results showed that the lesional hemoglobin was 2.8 times greater than that of adjacent uninvolved skin (Ryatt, Feather et al., 1983). Researchers have used NIRS to diagnose acute maxillary rhinosinusitis (Mansfield, Attas et al., 2005) and characterize the synovium of the small finger joints in both early and late rheumatoid arthritis (Canvin, Bernatsky et al., 2003). However, previous studies using NIRS have just shown relative changes in tissue oxygenation, hydration or blood volume (i.e. increase or decrease).

This study investigates the use of NIRS to assess Kaposi's sarcoma (KS) lesions. KS is a highly vascular tumor that is a frequent cause of morbidity and mortality among people infected with AIDS (Yarchoan, Tosato et al., 2005). Angiogenesis and capillary permeability play important roles in the development and progression of KS (Cornali, Zietz et al., 1996; Redondo, Sanchez-Carpintero et al., 2000). The most distinguishing visual characteristic of KS lesions is their color, which typically ranges from light brown to deep purple, with the latter color being the characteristic color for active lesions. Some of that color may be attributed to increased blood volume in the lesion. Visually inspecting, measuring and palpating skin lesions have long been used to assess the course of cutaneous disease in patients with KS. However, reliable assessment of these parameters requires a highly trained evaluator and evaluations made by different observers or even by the same observer at different times can be inconsistent. Quantitative instrumental methods offer a potentially more objective means of assessing skin and lesion characteristics to supplement the visual clinical observations. NIR spectroscopy has the potential to add functional information about the tumor and tissue oxygenation and blood volume that may aid in assessing the pathogenesis of the status and changes of the tumors during drug therapy.

Moreover, this approach may provide early markers for tumor responses to treatment and allow one to learn more about the pathophysiology of the disease.

MATERIALS AND METHODS

1. Instrumentation

The multi-spectral imager used in this work contains a cooled charge-coupled device (CCD) (CCD-612-TKB, Roper Scientific, Trenton, NJ) with spatial resolution 512 by 512 pixels and each image comprises a 16-bit digital signal. This array detector captures sequentially six images after the image forming light is passed through a set of narrow-band interference filters (having bandwidth of 40 nm at FWHM) centered at 700, 750, 800, 850, 900, and 1000 nm. This filter set is located on a computer-controlled rotating filter wheel positioned in front of the camera lens used to project the image into the CCD. The schematic diagram of the imaging system is shown in Figure 1a. The tissue is illuminated by a white light source (Dolan-Jenner Model 180, Edmund Optics, Barrington, NJ) which uses a 150 W EKE quartz halogen lamp. The illumination light is transported from the control unit via a 0.5 in-diameter optical fiber bundle. Polarization discrimination elements were used in order to remove the specular reflection component (which contains no functional information) from the recorded image (Demos and Alfano, 1997). Specifically, a first linear polarizer was placed at the output of the light transporting fiber bundle after passing through a red filter (550 nm long wavelength pass) (FS40-NIR-I, CVI Laser, Albuquerque, NM) used to protect the polarizer from excessive heating. A second linear polarizer was positioned in front of the CCD camera with its polarization orientation orthogonal to that of the illumination. Both polarizers were sheet polarizers for wavelengths 740-2400 nm (Thermo Oriel (now Newport), Stratford, CT, Part #27361). Since

initially constructing the system, changes have been made in the illumination assembly to make calibration easier. The schematic of the new instrumentation is shown in Figure 1b. In the upgraded system, the light transporting fiber bundle was terminated with an o-ring arrangement of the fibers to achieve nearly perpendicular incidence of the light with respect to the tissue surface from the approximately 45° angle in the original arrangement..

The system is operated using a personal computer running the Windows operating system (Microsoft, Seattle, WA) and incorporates a frame grabber card as the data acquisition hardware interface. Software has been developed to control the instrument driver routines for the CCD detector. The in-house software was written using LabVIEW (National Instruments, Austin, TX), an icon-based programming language designed for instrument control applications. The software allows the user to select experimental parameters such as the CCD detector shutter exposure time. An example of the six images collected by the CCD detector is shown in the top row of Figure 2. The grayscale bar shows the intensity of the reflected light, with white being a higher intensity. Images collected from the original and upgraded systems are collected and calibrated for the light intensity, light source and camera, as described in the following section.

2. Calibrations and Curvature Correction

The images collected from the system were calibrated to remove all elements of bias, including light intensity variations in space, wavelength-dependent camera sensitivity, and wavelength-dependent light intensity variations of the source. First, the light intensity variations in space were calibrated by normalizing the measured reflectance from the sample to the reflectance collected from a standard reference card (Eastman Kodak, Rochester, NY). This standard reference card provides 90% reflectance for unpolarized illumination and measurement

conditions. However, it is not characterized for polarization sensitive images, like we used in our experiments. For this reason, we only utilized the spatial distribution of the intensity in the cross polarization image of the reference card (which was normalized assuming an average peak intensity equal to 1) to correct the images of our samples for the non-uniform illumination conditions. However, we did not use these images as an absolute calibration standard. For this reason, our calibration procedure involved two additional steps.

The next calibration step involved the correction for the wavelength-dependent sensitivity of the camera. A standard source of spectral irradiance between 250-2400 nm (Oriel Part 63358, Spectra Physics Oriel, Stratford, CT) was used to determine the camera calibration curve including the CCD response and the optical elements attached. These values were then divided by the corresponding value at 750 nm resulting in a normalization factor of 1 at 750 nm. The results are summarized in Table 1. The high value in Table 1 at 700 nm (2.3) is mainly due to the lower transmission of the polarizer at this wavelength. On the other hand, the high value at 1000 nm (3.8) is due to the low sensitivity of the CCD. After each set of spectral images were normalized for the spatial variation of the illumination, they were then multiplied by the corresponding wavelength-dependent camera calibration factor.

Finally, the images have undergone a third normalization process to account for the wavelength-dependent light intensity variations of the illumination source. Images of the 90% reflectance card (also used in the first calibration step) after they were corrected for the spectral response of the camera were then used to determine the average intensity at 700, 750, 800, 850, 900, and 1000 nm to extract the normalization factor for each wavelength (which is also normalized to 1 at 750 nm as shown in Table 1). The imaging system provides a field-of-view of 25 cm² (3.88 in²). This differs from other approaches where point-like detectors are used to

acquire information at multiple locations. However, such an image can be analyzed to individual pixels or sets of pixels and thus treated as equivalent to a multi-source, multi-detector system. A set of initial and three step calibration-corrected images are shown in Figure 2.

A major challenge when collecting images from human tissue including extremities such as arms and legs is the natural curvature. In order to remove the curvature effects in the multi-spectral images, we have developed a curvature removal algorithm. After the three calibrations are performed on a set of images, the image is rotated to fit the primary curvature to the x- and y-axes. Then the average intensity for each column and each row is calculated, creating vertical and horizontal intensity curves, respectively. The vertical and horizontal intensity curves for each wavelength are then normalized and averaged over the utilized wavelengths. Then each pixel (x,y) in the image is divided by the normalized values of the corresponding x and y intensity curves. This allows the effects of the curvature in the horizontal and vertical direction to be removed, without changing the underlying hemodynamic effects.

This correction assumes the primary effect in intensity comes from the curvature and averaging will simply remove the fluctuations caused by hemodynamic change. Averaging the utilized wavelengths should also improve the model as any changes in intensity due to geometry should be wavelength independent. Visible lesion areas are disregarded for these calculations as they have intrinsically lower intensity than the surrounding tissue. We note that in patients with non-visible tumors, including the tumor in the analysis did not affect the results strongly as the primary intensity effect is once again curvature.

3. Mathematical Optical Skin Model

This research uses a mathematical optical skin model to determine the physiological tissue parameters of interest through a best-fit multivariate analysis. The light passes through the epidermis layer, diffusely reflects through the dermis layer, passes again through the epidermis layer, and is detected by the CCD camera. This yields the following general mathematical optical skin model equation:

$$I_{\text{det}}(\lambda) * I_{\text{calibration}}(\lambda) = S * A_{\text{epi}}(\lambda)^2 * A_{\text{dermis}}(\lambda) \quad (1)$$

where λ is wavelength (i.e. 700, 750, 800, 850, 900, and 1000 nm), $I_{\text{det}}(\lambda)$ is the detected intensity at each pixel for each wavelength, $I_{\text{calibration}}(\lambda)$ is the combined calibration factor described above for each wavelength, S is the scaling factor (to be discussed in section 3.3), $A_{\text{epi}}(\lambda)$ is the attenuation of the light intensity after passing through the epidermis for each wavelength, and $A_{\text{dermis}}(\lambda)$ is the attenuation of the light intensity after diffusely reflecting through the dermis layer for each wavelength. The epidermis takes melanin into account and the dermis includes the contributions from blood volume and blood oxygenation.

3.1 Epidermis Layer

The effect of the thin epidermis layer on the intensity of the diffusely reflected light is based on the volume fraction of melanin in the epidermis and the epidermal thickness (Jacques, Roman et al., 2000). The effective attenuation of light through the epidermis, A_{epi} , can be determined by Beer's Law giving:

$$A_{\text{epi}}(\lambda) = e^{-\mu_{a(\text{epi})}(\lambda)t} \quad (2)$$

where λ is the wavelength [nm], $\mu_{a(epi)}(\lambda)$ is the epidermis absorption coefficient [mm^{-1}], and t is the thickness of the epidermis [mm]. The epidermis absorption coefficient, $\mu_{a(epi)}(\lambda)$, depends on the percent of melanin in the skin and is given by:

$$\mu_{a(epi)}(\lambda) = V_{mel}\mu_{a(mel)}(\lambda) + (1 - V_{mel})\mu_{a(skin)}(\lambda) \quad (3)$$

where V_{mel} is the volume fraction of melanin depending on the complexion of the skin, $\mu_{a(mel)}(\lambda)$ is the melanin absorption coefficient [mm^{-1}], and $\mu_{a(skin)}(\lambda)$ is the absorption coefficient of normal skin excluding melanin (baseline) [mm^{-1}]. For the results in this paper, previously reported values for $\mu_{a(mel)}(\lambda)$ (Meglinski and Matcher, 2002) and $\mu_{a(skin)}(\lambda)$ (Jacques, 1998) have been used (see Table 2). The remaining variable is the volume fraction of melanin in the epidermis, V_{mel} . The absorption coefficient for normal skin was considered the same for the epidermis and dermis layers. This model does not consider epidermal scattering because the probability of one scattering event in an epidermis of 0.06 mm is very small (less than 5%).

3.2 Dermis Layer

The influence of the much thicker, highly scattering dermis layer on the skin reflectance is estimated by a stochastic model of photon migration, e.g., random walk theory. Given the known random walk expression for diffuse reflectivity, R_o , of a semi-infinite medium with absorbing boundary and no index mismatch (Gandjbakhche, 1995):

$$R_o(\mu) = \frac{e^{-2\mu}}{\sqrt{24\mu}} [1 - \exp(-\sqrt{24\mu})] \quad (4)$$

where μ is the ratio of absorption and scattering coefficients $\left(\frac{\mu_a}{\mu'_s}\right)$. As shown by our Monte

Carlo simulations, the effect of index mismatch on diffuse reflectivity is reduced to an additional

factor K in equation (4). In the case of tissue (refractive index $(n) = 1.4$) / air boundary K equals approximately to 0.57. In our analysis, this constant factor is incorporated into the scaling factor, S , discussed in Section 3.3. To simplify computations, equation (4), as a function of μ , has been fitted by a power law plus constant. It was shown, that for reasonable values of optical parameters of tissues in the red/near-infrared spectral range ($5 \times 10^{-4} < \mu < 0.1$), equation (4) is well approximated by:

$$R_o(\mu) = 1.06 - 1.45\mu^{0.35}, \quad (5)$$

In case of the dermis, $\mu = \frac{\mu_{a(dermis)}(\lambda)}{\mu'_s(\lambda)}$ where $\mu'_s(\lambda)$ and $\mu_{a(dermis)}(\lambda)$ are the reduced scattering and absorption coefficients of the dermis, respectively. Plugging μ into equation (5) yields the equation for the dermis attenuation used in equation (1):

$$A_{dermis}(\lambda) = 1.06 - 1.45 \left[\frac{\mu_{a(dermis)}(\lambda)}{\mu'_s(\lambda)} \right]^{0.35} \quad (6)$$

The dermis absorption coefficient depends on the volume fraction of blood in the tissue (V_{blood}) and the percent of that blood that is oxygenated (V_{oxy}) (relative fractions of HbO₂ and Hb in the blood). The dermis absorption coefficient [mm^{-1}] is calculated by:

$$\mu_{a(dermis)}(\lambda) = V_{blood} \mu_{a(blood)}(\lambda) + (1 - V_{blood}) \mu_{a(skin)}(\lambda) \quad (7)$$

where V_{blood} is the volume fraction of blood in the dermis layer and $\mu_{a(skin)}(\lambda)$ is the absorption coefficient of skin [mm^{-1}]. The absorption coefficient of whole blood is calculated as [mm^{-1}]:

$$\mu_{a(blood)}(\lambda) = V_{oxy} \mu_{a(oxy)}(\lambda) + (1 - V_{oxy}) \mu_{a(deoxy)}(\lambda) \quad (8)$$

where V_{oxy} is the fraction of blood that is oxygenated, $\mu_{a(oxy)}(\lambda)$ is the absorption coefficient of HbO₂ and $\mu_{a(deoxy)}(\lambda)$ is the absorption coefficient of Hb. For $\mu_{a(oxy)}(\lambda)$ and $\mu_{a(deoxy)}(\lambda)$,

previously published values were used (Delpy, Cope et al., 1988). The reduced scattering coefficient, $\mu'_s(\lambda)$, is calculated by combining 40% Mie and 60% Rayleigh scattering components given by (Jacques, 1996):

$$\mu'_{s(mie)}(\lambda) = 2 \times 10^4 \lambda^{-1.5} \quad (9)$$

$$\mu'_{s(rayleigh)}(\lambda) = 2 \times 10^{11} \lambda^{-4} \quad (10)$$

These percentages for Mie and Rayleigh scattering components gave reasonable results when analyzing tissue from normal subjects. The wavelength dependence of the reduced scattering coefficient obtained from equations (9-10) is shown in Table 2. This combination represents the distribution of sizes of different types of scatterers in tissue, from granules to cells. Other analytes such as hemosiderin, iron deposits created after hemoglobin is degraded, may be added into the model in the future.

3.3 Analysis

The sampling volume, considered to be the area of tissue that affects the light absorption and reflectance, is larger than the individual pixel on the surface of the skin. Each image pixel is affected by the surrounding tissue and therefore the reconstructed images of blood volume and blood oxygenation are averages of the sampling volume and are influenced by the tissue surrounding that area. The sampling volume also depends on the penetration depth of the near-infrared light source.

Preliminary Monte-Carlo simulations of the multi-spectral system and skin model have been run to determine the maximum penetration depth. The volume visited by the photons depends on the wavelength. Moreover, ~2 mm seems to be the limit of the photon penetration for

the theoretical instrumental setup i.e., the photons mainly remain in the epidermal and dermal layer. It is interesting to notice that the voxels “visited” by the photons appear to be situated vertically and symmetrically under the detection pixel. In practice, this means that an image formed at the skin surface should carry information concerning the tissue mainly situated under each pixel.

Equation (1) also contains a scaling factor required to balance the two sides of the equation. This scale factor varies between and among subjects, but a standardized equation has been developed:

$$S = 2.2382 * \bar{I} \quad (12)$$

where S is the scaling factor and \bar{I} is the average intensity of a region of interest from the wavelengths of interest. The scale factor has a linear relationship with the calibrated intensity averaged across the four wavelengths ($R^2 = 0.8839$) (Figure 3). This equation was developed by analyzing normal subject data where reasonable values of tissue oxygenation and blood volume based on published data were calculated.

A nonlinear fit of the calibrated intensity of each pixel to the attenuation of light through both layers of the mathematical optical model is performed using the Levenberg-Marquardt algorithm. For each pixel, four equations are created, one per wavelength at 700, 750, 800 and 850 nm, of equation (1). By inserting equations (2, 3, 6-10) into equation (1), four unknown physiological parameters can be solved for, V_{oxy} , V_{blood} , V_{mel} and t .

With the multi-spectral imaging system we designed, we have the ability to use all six wavelength images and solve for two additional parameters. In future studies, when the 900 and 1000 nm images are used, one must account for the absorption of water and lipids in the model and consider their volumes as variables. For this research, t was considered to be 0.06 mm based

on previous research (Nishidate, Aizu et al., 2003). The value of V_{mel} was considered constant for each subject based on the range published by Jacques for different skin types (Jacques, 1996). Performing a best-fit analysis of this set of equations allows oxygenated hemoglobin and tissue blood volume to be solved for.

The spatial map of oxygenated hemoglobin obtained from this analysis presents the fraction of blood in a given volume of tissue that is oxygenated. The values can range from zero to one. For example, a value of 0.80 means that the blood in a specific area of tissue is 80% oxygenated. Similarly, the spatial map of blood volume shows the fraction of blood in the tissue sampling volume (it can also vary between zero and one). For example, a value of 0.05 means that the sampled volume of tissue contains 5% blood.

4. Laser Doppler Imaging (LDI)

This study combines multi-spectral imaging and laser Doppler imaging (LDI) to determine changes in blood volume, oxygenation state and blood velocity of the microvasculature. Using both imaging modalities provides a detailed analysis of tissue function that can be assessed as a possible predictive tool for the outcome and individualization of therapeutic strategies. LDI enables noninvasive analysis of blood flow patterns in skin up to an approximate depth of one millimeter. We used a dual wavelength (MoorLDI, Moor Instrument, Inc, UK) imager for simultaneous scanning at two wavelengths, 690 nm (visible red) and 780 nm (NIR) with spatial resolution of approximately 100 μm for assessing blood flow from differing microvascular beds (Wardell, Jakobsson et al., 1993). LDI produces a two-dimensional image of tissue blood perfusion by scanning a low power solid-state laser beam (5 mW Helium-Neon) over the tissue of interest. The light is scattered by moving red blood cells and LDI

measures the Doppler shift due to this movement. The output of LDI commonly used to describe blood flow measurement is ‘flux’: a quantity proportional to the product of the average speed of the blood cells (often referred to as blood velocity) and their number concentration (referred to as blood volume). LDI more directly measures the blood perfusion of small blood vessels in tissue, which generally increases as the blood supply increases during angiogenesis (Sorensen, Bengtsson et al., 1996; Rivard, Fabre et al., 1999).

RESULTS AND DISCUSSION

1. Normal Subjects

To quantify possible changes in physiological parameters due to disease presence/progression, a preliminary study of healthy subjects is required. Twelve healthy subjects were studied during the following experiments. One of the first experiments we conducted compared a similar area on both arms of the same healthy subject (Figures 4 and 5). The first column of both figures shows two 700 nm images collected from the multi-spectral imaging system of the left and right forearm of the same subject. The second and third columns show the reconstructed images presenting the fractions of V_{oxy} and V_{blood} , respectively, of similar areas on the left and right forearm of one subject. Observed values of oxygenated hemoglobin in both figures are lower than 0.6, corresponding to blood oxygenation below 60%. Similarly, values of blood volume are lower than 0.2, corresponding to blood of less than 20% per sampling volume. Both figures show an average value of 0.07 for the background skin blood volume, meaning that each tissue sampling volume contains an average of 7% blood, is close to agreement with the range given by Jacques of 0.2-5% (1996). Interestingly, a higher blood volume was observed in the vasculature compared to the surrounding tissue. The blood vessels

detected by this technique are veins and arteries. Also, the HbO_2 images show a higher percentage of oxygenated hemoglobin in the vasculature compared to the surrounding tissue, showing an average of approximately 47% oxygenated hemoglobin in the sampled tissue volume. The images also show consistent percentages of oxygenated hemoglobin and volume of blood in the vasculature and surrounding tissue on both arms.

The next study followed healthy subjects through reactive hyperemia. A pressure cuff was inflated around the subject's upper arm to induce temporary ischemia, which was relieved after several minutes. Spectroscopic image sets were taken every 30 seconds at different time points: before the occlusion (0 min), during the occlusion (0.5-5 min), and after the occlusion was released (5.5-10 min). An example of these changes is presented in Figure 6 at three time points: before the occlusion, after five minutes of occlusion, and one minute after the occlusion was released. The pairs of spectroscopic images illustrate the changes in the distributions of oxygenated hemoglobin and blood volume fractions, reconstructed from these data sets. The tissue around the vasculature appears substantially more deoxygenated during the occlusion episode than afterwards. In the bottom row of Figure 6, the skin shows a slight increase in background blood volume during the occlusion period. This follows previous results showing that during an occlusion, there will be an increase in blood volume and a decrease in oxygenated hemoglobin (Kooijman, Hopman et al., 1997; Kragelj, Jarm et al., 2000; Kragelj, Jarm et al., 2001). When the occlusion is released, the oxygenated hemoglobin increases past the baseline and then returns to the baseline following several minutes.

Additional reactive hyperemia experiments were conducted on healthy subjects in a more longitudinal manner. Figure 7 shows the results of two visits by one healthy subject. Similar patterns of changes in V_{oxy} and V_{blood} during the experiments can be seen. In addition, similar

values of initial blood volume and oxygenated hemoglobin in the vasculature and surrounding tissue were observed. This experiment helps to show the stability of the system and the reconstruction algorithm. In general, spectroscopic images allow us to quantify circulatory changes of the skin in time to study physiological variations in different tissue locations.

2. Kaposi's Sarcoma Subjects

To study diseased tissue, clinical protocols were approved by the institutional review board of the National Cancer Institute to study subjects with Kaposi's sarcoma lesions. Written consent was obtained from all subjects and lesions were considered for measurement if they were at least 0.2 cm in diameter. Immediately prior to imaging, each patient removed sufficient garments and accessories to expose the entire area to be recorded including the contralateral side. Prior to and during imaging, the subjects were seated at rest in a closed room around 23°C.

Figure 8 shows a set of comparative images from one visit of a Kaposi's sarcoma subject. The digital image in the top left shows the visible location of the surface lesion; however, the non-invasive imaging techniques show additional information about the lesion and its vasculature. To allow the visualization of the lesion in each modality image, a black line marks the location of the lesion based on the digital camera image. One should note the presence of a vasculature network surrounding the tumor. The average oxygenation of veins in the bottom left image corresponds to the known value of approximately 47%. There is increased blood flux (top right), blood volume (bottom right) and fraction of HbO₂ (bottom left) in the lesion compared to the surrounding tissue.

To date, over 30 patients have been studied at entry onto treatment protocols of several experimental therapies. The overall clinical response of KS to this treatment regimen was assessed using a minor modification of the AIDS Clinical Treatment Group (ACTG) KS

parameters as previously published (Krown, Testa et al., 1997; Little, Wyvill et al., 2000). In short, responses were based on changes in the total number of KS lesions, the number of nodular KS lesions, the sum of the products of the larger perpendicular diameters of five representative “marker” lesions selected at entry onto the protocol, tumor associated edema or effusions, and visceral disease. A complete response was defined as the absence of all detectable disease for at least four weeks with a biopsy of at least one previous representative lesion showing the absence of malignant cells. Many patients that did not experience a positive response on one therapy tried a different one.

Figure 9 shows a set of comparative images of a KS patient undergoing an experimental drug therapy. It is important to note that as this patient returns for subsequent visits, the clinical assessment was stable disease. After three weeks of experimental treatment, increased blood flux and blood volume can be seen in the lesion compared to the surrounding tissue.

A set of comparative images of a KS patient collected at baseline (0 weeks) and at follow-up (19 weeks later) is shown in Figure 10. The clinical assessment of this patient, who was receiving retroviral therapy only (no experimental drug therapy), was progressive disease by week 13. The images in the left column show increased HbO_2 and tissue blood volume fractions and blood flux values in the lesion compared to the background at baseline. Note that a ballpoint pen was used to write a number next to the lesion (here a 4). This ink appears to have low HbO_2 values; however, the ink pixels were not used to calibrate the image, but the ink has strong absorption properties and behaves differently than tissue in the reconstruction. It is left in the image as a visual reference for the clinician, but its values can be disregarded. Nineteen weeks later, the images in the right column show a lower HbO_2 fraction in the lesion compared to the background, but continue to show increased tissue blood fraction and blood flux in the lesion

area compared to the background tissue. The laser Doppler images show increased blood circulation in the entire viewing area after 19 weeks. The tissue blood volume fractions were both approximately 6% in the tissue surrounding the lesion. A square region of interest (ROI) was selected inside the lesion and compared to the reference region outside the lesion. A ROI of 25x25 pixels (625 pixels²) was chosen inside the lesion (unless the lesion was smaller than this and then an area of 10x10 was chosen). The four 25x25 pixel ROIs chosen as the reference regions and averaged together are shown in the last row of Figure 10. Parameter averages were calculated for both areas. Table 3 shows the quantitative results of tissue HbO₂ and blood volume fractions from the subject in Figure 10.

Results from this patient shows that as the disease progressed, the lesion became more hypoxic and the surrounding tissue became more oxygenated. Blood volume was fairly constant as the disease progressed. The blood flux, as shown in the laser Doppler images, shows a slight decrease inside the lesion as the disease progressed, but an increase of blood circulation in the area surrounding the lesion. We previously found a strong correlation between temperature, measured by thermal imaging, and blood velocity, as well as with blood volume (Hassan, Little et al., 2004). Further analysis of the data we are collecting promises to further our understanding of tumor angiogenesis and the effects of experimental therapies by using multi-spectral imaging, LDI and thermal imaging.

CONCLUSIONS

The multi-spectral imaging system reconstructs images with contrast based on spatial changes in the concentrations of analytes. The results show that quantitative spectroscopic imaging can be used for dermatological examinations. Objective information on skin health is obtainable non-

invasively and without skin contact, up to a depth of approximately 2 mm. Quantitative tissue oxygenation and blood volume data can be provided independently, with detailed indications of variability with location, in the form of images.

The quantitative reconstructed values, or contrast, have shown utility in revealing spatial patterns of analyte concentration that correlate with KS tumor regions and can be used to show informative changes indicative of patient response to therapy. Consistent with their angiogenic nature, the KS lesions studied had greater blood volume than surrounding normal tissues. However, some appeared to be relatively hypoxic compared to the surrounding tissue, while others appeared to have increased blood oxygenation. KS involves a degree of inflammation, and it is possible that the areas of increased oxygenation reflect inflammation. It will be interesting to correlate these variations with other parameters of the KS lesions in future analyses.

It should be noted that these techniques involve modeling based on certain assumptions and calculations, and it may be possible to refine the techniques further as we proceed. For example, the model now does not include subcutaneous hemosiderin, and including this parameter may yield improved methodology, especially in KS lesions.

Overall, multi-spectral imaging is a non-invasive technique that can detect the progression of disease before these changes are visibly apparent. This technique is now also being considered for use in studies of other skin characteristics, including erythema, blanching, edema, and fibrosis. Near-infrared spectroscopic imaging has potential to be the technique of choice to supplement visual examination.

ACKNOWLEDGEMENTS

This research was funded in part by the Intramural Research Program of the National Institute of Child Health and Human Development and the National Cancer Institute of the National Institutes of Health. This work was performed in part at UC, Lawrence Livermore National Laboratory under the auspices of the U.S. Department of Energy under Contract W-7405-Eng-48. The authors would also like to thank Kathleen Wyvill and Karen Aleman who assisted with the KS patient visits and Israel Gannot for his numerous useful discussions.

REFERENCES

- Attas, M., M. Hewko, et al. (2001). "Visualization of cutaneous hemoglobin oxygenation and skin hydration using near-infrared spectroscopic imaging." Skin Research and Technology **7**: 238-245.
- Canvin, J. M. G., S. Bernatsky, et al. (2003). "Infrared spectroscopy: shedding light on synovitis in patients with rheumatoid arthritis." Rheumatology **42**: 76-82.
- Cornali, E., C. Zietz, et al. (1996). "Vascular endothelial growth factor regulates angiogenesis and vascular permeability in Kaposi's sarcoma." Am J Pathol **149**(6): 1851-69.
- Delpy, D. T., M. Cope, et al. (1988). "Estimation of Optical Pathlength through Tissue from Direct Time of Flight Measurement." Physics in Medicine and Biology **33**(12): 1433-1442.
- Demos, S. and R. Alfano (1997). "Optical Polarization Imaging." Applied Optics **36**: 150-155.
- Gandjbakhche, A. H., Weiss, G. H. (1995). "Random walk and diffusion-like models of photon migration in turbid media." Progress in Optics **XXXIV**(V): 335-402.

- Hassan, M., R. F. Little, et al. (2004). "Quantitative Assessment of Tumor Vasculature and Response to Therapy in Kaposi's Sarcoma Using Functional Noninvasive Imaging." Technol. Cancer Res. Treat. **3**(5): 451-457.
- Jacques, S., J. Roman, et al. (2000). "Imaging superficial tissues with polarized light." Lasers in Surgery and Medicine **26**(2): 119-29.
- Jacques, S. L. (1996). Origins of Tissue Optical Properties in the UVA, Visible, and NIR Regions. OSA TOPS on Advances in Optical Imaging and Photon Migration, OSA.
- Jacques, S. L. (1998). "Skin Optics." from <http://omlc.ogi.edu/news/jan98/skinoptics.html>.
- Kollias, N., R. Gillies, et al. (1995). "A single parameter, oxygenated hemoglobin, can be used to quantify experimental irritant-induced inflammation." Journal of Investigative Dermatology **104**(3): 421-424.
- Kooijman, H. M., M. T. E. Hopman, et al. (1997). "Near Infrared Spectroscopy for Noninvasive Assessment of Claudication." Journal of Surgical Research **72**: 1-7.
- Kragelj, R., T. Jarm, et al. (2001). "Parameters of Postocclusive Reactive Hyperemia Measured by Near Infrared Spectroscopy in Patients with Peripheral Vascular Disease and in Healthy Volunteers." Annals of Biomedical Engineering **29**: 311-320.
- Kragelj, R., T. Jarm, et al. (2000). "Reproducibility of Parameters of Postocclusive Reactive Hyperemia Measured by Near Infrared Spectroscopy and Transcutaneous Oximetry." Annals of Biomedical Engineering **28**: 168-173.
- Krown, S. E., M. A. Testa, et al. (1997). "AIDS-related Kaposi's sarcoma: prospective validation of the AIDS Clinical Trials Group staging classification. AIDS Clinical Trials Group Oncology Committee." J Clin Oncol **15**(9): 3085-92.

- Little, R. F., K. M. Wyvill, et al. (2000). "Activity of thalidomide in AIDS-related Kaposi's sarcoma." J Clin Oncol **18**(13): 2593-602.
- Mansfield, C., E. Attas, et al. (2005). "Evaluation of static thermal and near-infrared hyperspectral imaging for the diagnosis of acute maxillary rhinosinusitis." J Otolaryngol **34**(2): 99-108.
- Meglinski, I. V. and S. J. Matcher (2002). "Quantitative assessment of skin layers absorption and skin reflectance spectra simulation in the visible and near-infrared spectral regions." Physiol. Meas. **23**: 743-753.
- Nishidate, I., Y. Aizu, et al. (2003). "Estimation of Absorbing Components in a Local Layer Embedded in the Turbid Media on the Basis of Visible and Near-Infrared (VIS-NIR) Reflectance Spectra." Optical Review **10**(5): 427-435.
- Redondo, P., I. Sanchez-Carpintero, et al. (2000). "Classic Kaposi's sarcoma and vascular endothelial growth factor." Acta Derm Venereol **80**(3): 218-9.
- Rivard, A., J. E. Fabre, et al. (1999). "Age-dependent impairment of angiogenesis." Circulation **99**(1): 111-20.
- Ryatt, K., J. Feather, et al. (1983). "The usefulness of reflectance spectrophotometric measurements during psoralens and ultraviolet A therapy for psoriasis." J Amer Acad Derm **9**(4): 558-562.
- Shah, N., A. Cerussi, et al. (2001). "Noninvasive functional optical spectroscopy of human breast tissue." Proc Natl Acad Sci USA **98**(8): 4420-4425.
- Sorensen, J., M. Bengtsson, et al. (1996). "Laser Doppler perfusion imager (LDPI)--for the assessment of skin blood flow changes following sympathetic blocks." Acta Anaesthesiol Scand **40**(9): 1145-8.

- Stranc, M. F., M. G. Sowa, et al. (1998). "Assessment of tissue viability using near-infrared spectroscopy." British Journal of Plastic Surgery **51**: 210-217.
- Svaasand, L. O., L. T. Norvang, et al. (1995). "Tissue Parameters Determining the Visual Appearance of Normal Skin and Port-wine Stains." Lasers in Medical Science **10**: 55-65.
- Taylor, D. and S. Simonson (1996). "Use of near-infrared spectroscopy to monitor tissue oxygenation." New Horiz. **4**(4): 420-425.
- Wagnieres, G., S. Cheng, et al. (1997). "An optical phantom with tissue-like properties in the visible for use in PDT and fluorescence spectroscopy." Phy. Med. Biol **42**: 1415-1426.
- Wardell, K., A. Jakobsson, et al. (1993). "Laser Doppler perfusion imaging by dynamic light scattering." IEEE Trans Biomed Eng **40**(4): 309-16.
- Yarchoan, R., G. Tosato, et al. (2005). "Therapy insight: AIDS-related malignancies - the influence of antiviral therapy on pathogenesis and management." Nature Clin Prac Oncology **2**(8): 406-15.

Tables

Table 1. Source and camera calibration factors by wavelength

Wavelength (nm)	Approximate Source Calibration Factor	Approximate Camera Calibration Factor	Approximate Source and Camera Calibration Factor
700	1.0	2.3	2.3
750	1.0	1.0	1.0
800	1.1	0.89	0.98
850	1.4	0.86	1.2
900	1.4	1.2	1.7
1000	2.3	3.8	8.7

Table 2. Optical properties used in the mathematical skin model

Wavelength (nm)	$\mu_{a(deoxy)}$ [1/mm]	$\mu_{a(oxy)}$ [1/mm]	$\mu_{a(skin)}$ [1/mm]	$\mu_{a(mel)}$ [1/mm]	μ'_s [1/mm]
700	1.07	0.23	0.0266	1.68	0.93
750	0.79	0.33	0.0254	1.33	0.77
800	0.48	0.47	0.0249	1.08	0.65
850	0.43	0.58	0.0246	0.88	0.55

Table 3. Quantitative results from reconstructed NIR spectroscopy images of a KS subject with progressive disease of oxygenated hemoglobin and blood volume.

	HbO₂ Fraction		Blood Fraction		Laser Doppler	
	<i>Baseline</i>	<i>19 weeks</i>	<i>Baseline</i>	<i>19 weeks</i>	<i>Baseline</i>	<i>19 weeks</i>
Lesion (ROI mean)	0.38	0.27	0.17	0.15	830	769
Outside (ROI mean)	0.22	0.29	0.07	0.06	137	372

Figure Captions

Figure 1. (a) Schematic of the original NIR multi-spectral imaging system. The filter wheel in front of the CCD camera contains six filters of 700, 750, 800, 850, 900, and 1000 nm. (b) Schematic of upgraded NIR spectroscopy system. The light source is projected directly in front of the CCD camera so that the light would shine perpendicular to the object surface.

Figure 2. Spectral images of a normal subject's arm at six NIR wavelengths (top row) before any calibrations and (bottom row) after intensity, source and camera calibrations. The grayscale bar shows the intensity of the reflected light. Note that tissue image intensity is normalized after the calibrations.

Figure 3. Fit of the standardized scaling factor that is used in Equation (1) ($R^2 = 0.8839$). This factor is different for each person and each visit. Each point on the graph indicates a normal subject with reasonable values of tissue oxygenation and blood volume based on published data.

Figure 4. Images comparing a normal subject's left and right forearm. Column 1 shows the calibrated 700 nm multi-spectral images collected from the system of each forearm. Column 2 shows the spatial maps of blood volume (V_{blood}) and oxygenated hemoglobin (V_{oxy}) were reconstructed from the set of NIR spectral images. Note that both arms show similar quantitative values for oxygenated hemoglobin and blood volume.

Figure 5. Images comparing a normal subject's left and right forearm. Column 1 shows the calibrated 700 nm multi-spectral images collected from the system of each forearm. Column 2 shows the spatial maps of blood volume (V_{blood}) and oxygenated hemoglobin (V_{oxy}) were reconstructed from the set of NIR spectral images. Note that both arms show similar quantitative values for oxygenated hemoglobin and blood volume.

Figure 6. Experiment showing (top row) oxygenated hemoglobin and (bottom row) blood volume during occlusion showing reactive hyperemia. (a) 0 mins, (b) after 5 mins of occlusion, and (c) 1 min after releasing occlusion. During the occlusion, V_{oxy} decreased and then when the occlusion was released, V_{oxy} increased above the initial state. A slight increase in V_{blood} was seen after 5 min occlusion period.

Figure 7. Experiment showing oxygenated hemoglobin and blood volume during occlusion showing reactive hyperemia for the same subject on two visits (top row is visit 1 and bottom row is visit 2). 1 (bottom row) Visit 2 (a) V_{oxy} at 0 mins, (b) V_{oxy} after 5 mins of occlusion, (c) V_{oxy} 1 min after releasing occlusion, (d) V_{blood} at 0 mins, (e) V_{blood} after 5 mins of occlusion, and (f) V_{blood} 1 min after releasing occlusion. During the occlusion, V_{oxy} decreased and then when the occlusion was released, V_{oxy} increased above the initial state. A slight increase in V_{blood} was seen after 5 min occlusion period and decreased toward the initial state 1 min after releasing the occlusion.

Figure 8. Images of Kaposi's sarcoma subject on entry onto drug treatment protocol. (top left) digital image; (top right) laser Doppler image showing blood flux; (bottom left) reconstructed map of fraction of oxygenated hemoglobin; (bottom right) reconstructed map of fraction of blood in tissue. There is increased blood flux, blood volume and fraction of HbO₂ in the lesion compared to the surrounding tissue.

Figure 9. Images of Kaposi's sarcoma subject (top row) on entry onto drug treatment protocol and (bottom row) after three weeks of experimental therapy. (a) digital image; (b) laser Doppler image showing blood flux; (c) reconstructed map of fraction of oxygenated hemoglobin; (d) reconstructed map of fraction of blood in tissue. The lesions appear hypoxic compared to the surrounding, but the laser Doppler and blood volume images show higher blood flux and volume in the lesions.

Figure 10. Images of Kaposi's sarcoma subject receiving antiretroviral therapy (left column) upon entry onto protocol at 0 weeks and (right column) follow-up 19 weeks later. (top) digital image; (row one) reconstructed map of oxygenated hemoglobin; (row two) reconstructed map of blood volume; (row three) laser Doppler images; (row 4) 700 nm spectroscopic image showing the four ROIs used to average the oxygenated hemoglobin and blood volume outside the lesion. This patient had a clinical diagnosis of progressive disease before the follow-up visit. Decreased oxygenated hemoglobin fraction, blood volume and blood flux in the lesion are apparent at the follow-up visit. Increased overall vasculature and oxygenated hemoglobin can be seen in the laser Doppler and oxygenated hemoglobin images at follow-up.